



Certificate of Conformity

VWR MEASURING CYLINDER, CLASS A, 100 ml

VWR International certifies that the specified products meet all VWR specifications and quality criteria for release. The products comply in every respect with the relevant particular specifications, drawings, relevant standards and regulations in force.

VWR International certifie que les produits mentionnés respectent toutes les spécifications et critères de qualité VWR. Les produits satisfont aux spécifications particulières, aux plans, aux normes et règlements en vigueur s'y rapportant.

VWR International bestätigt hiermit, daß die angegebenen Produkte alle VWR Spezifikationen und Qualitätskriterien zur Freigabe erfüllen. Die Produkte entsprechen in jeder Hinsicht den maßgeblichen Spezifikationen, den Zeichnungen sowie den relevanten Normen und gültigen Vorschriften.

Cat. No.	Lot No.	Capacity ml	Division ml	Tolerance	Pack
612-3836	04.12	100	1	0.5 $\pm$ ml	2

Material : Borosilicate glass, tall form

QC Testing: Representative production samples are collected and inspected in accordance with current applicable product specifications.

This document has been produced electronically and is valid without a signature

Date of issue: **December 2012**

VWR International bvba/sprl, Researchpark Haasrode Zone 2020, Geldenaaksebaan 464, 3001 Leuven, Belgium
<http://www.vwr.com>



Version: 2.0