Certificate of Conformity

VWR CELL CULTURE DISH, STANDARD LINE

VWR International certifies that the specified products meet all VWR specifications and quality criteria for release.The products comply in every respect with the relevant particular specifications, drawings, relevant standards and regulations in force.

VWR International certifie que les produits mentionés respectent toutes les specifications et critères de qualité VWR. Les produits satisfont aux spécifications particulières, aux plans, aux normes et règlements en vigueur s'y rapportant.

VWR International bestätigt hiermit, daß die angegebenen Produkte alle VWR Spezifikationen und Qualitätskriterien zur Freigabe erfüllen. Die Produkte entsprechen in jeder Hinsicht den maßgeblichen Spezifikationen, den Zeichnungen sowie den relevanten Normen und gültigen Vorschriften.

VWR Catalog No: **734-2322**

Lot No: **110510-116**

Growth area cm²	Ø	Height	Packed	Pk
143.0	**150 mm**	**22 mm**	**1/bag**	**120**

Material: **100% High Clarity Polystyrene**

QC Testing: **Representative production samples are collected and inspected in accordance with current applicable product specifications.**

Cytotoxity: **Testing is conducted to quality all material resings using ISO 10993 standards for cytotoxicity and have been shown to be non-toxic.**

Non-pyrogenic: **The acceptance level of product is 0.5 EU/ml or less than 20 EU/device. LAL Gel Clot Method**

Sterility: **Products labeled sterile meets a minimum requirement of 10^{-6} SAL.**

Date of Sterilization: **14th May 2011**

Expiry Date: **14th May 2014**

This document has been produced electronically and is valid without a signature

Date of Issue: **January 2013**