Certificate Of Conformity

VWR BLENDER BAGS WITH FULL FILTER STERILE

VWR International certifies that the products specified conform to all specifications and quality criteria mentioned in this document.

VWR International certifie que les produits mentionnés respectent toutes les specifications et critères de qualité mentionnés dans ce document.

VWR International bestätigt hiermit, daß die angegebenen Produkte alle in diesem Dokument erwähnten Spezifikationen und Qualitätskriterien erfüllen.

Cat. No.	Description	Capacity (ml)	Height (mm)	Width (mm)	Thickness (mm)	Pk
129-1106	Blender bags with full filter sterile	400	300±3	190±3	0,072	10 bags of 50 pcs

Material: multicoated reinforced complex
Full-page calibrated microperforated filter
Filter porosity: 280 microns

Food contact : Approve for food contact

Manufactured under ISO 9001:2008

Sterilization : Is made in compliance with to the GMP of quality Assurance and attest treatment by gamma rays between 5 to 12 Kilo Gray

Blender test : Bags have to be resistant for 3 minutes blending (frequency: 4 times / second)

Use temperature main body : 40°C / +80°C

Expiry: 10 years

Physical Requirements
Flexible, good mechanical resistance.
Can be frozen
Red sticker on every box serves as qualitive indicator of gamma ray processing.

Disclaimer: *VWR wants to state that these specifications can only be regarded as general information and will not discharge the user from his obligation to ensure the product is suitable for the intended use - If there is any doubt ask us.*

Date of issue: **November 2015**