



Certificate of Conformity

VWR TWIST-SEAL BAGS CLOSURE WITH 1 ROUND AND 1 FLAT WIRE, WITH WRITE-ON PANEL

VWR International certifies that the products specified conform to all specifications and quality criteria mentioned in this document.

VWR International certifie que les produits mentionnés respectent toutes les spécifications et critères de qualité mentionnés dans ce document.

VWR International bestätigt hiermit, daß die angegebenen Produkte alle in diesem Dokument erwähnten Spezifikationen und Qualitätskriterien erfüllen.

Cat. No.	Capacity (ml)	Thickness (µm)	Tabs	Dimensions (mm)	Pk
129-9020	150	63	Safety	76×178	500
129-9860	450	63	Regular	114×229	500
129-9862	650	76	Regular	140×229	500
129-9861	800	63	Regular	114×305	500
129-9863	1650	76	Regular	178×305	250
129-9888	450	63	Safety	114×229	500
300-0106	650	76	Safety	140×229	1.000
300-0107	1650	76	Safety	178×305	1.000
129-0012	2250	101	Safety	254×305	250
129-0013	4000	101	Safety	254×382	250

Bag material : Virgin polyethylene tubing (FDA approved)

Closure material : Aluminium or steel wires covered with FDA approved pressure sensitive tape with a backing of polypropylène or PVC

- **Closure type :** 1 round, 1 Flat wires

Sterility :

Sterility is assured due to extrusion temperatures exceeding 220°C which eliminates any present bacteria. During manufacturing all bags are sealed at both top and bottom in a continuous format before being cut into individual bags. This method assures that the bag's interior will never come in contact with its surroundings. Periodic validation tests, based on USP 34 – Sterility testing, are performed in our facility to guaranty the sterility of our processes.



Certificate of Conformity

Compliant to: FDA

This is to confirm that the LDPE used in the production of our sterile sampling bags conforms to the applicable regulations to the Food and Drug Administration [21 CFR 177.1520 (c)3.1a, (c)3.2a and 178.2010] for use in contact with nonalcoholic food, including use in articles used for packing and holding food during cooking.

Other characteristics:

- Provide a secure, contaminant-free pliant container that ensures dependable analysis results
- Offer an economical and efficient way to collect, contain and carry samples
- Wire tab keeps bag open for filling
- Unique fold over system providing a temporary leak-proof seal
- Sterile material
- R-Nase, D-Nase and Pyrogene free

Disclaimer: VWR wants to state that this declaration will not discharge the user from his obligation to ensure the product is suitable for the intended use. - If there is any doubt ask us.

Date of issue: **February 2016**