Certificate of Conformity

VWR DEEP WELL PLATES, 96-WELL, NOTCHED CORNERS

VWR International certifies that the products comply in every respect with the relevant particular specifications and/or drawings and/or relevant standards and regulations in force, mentioned in this document.

VWR International certifie que les produits sont conformes à tous égards aux spécifications particulières appropriées et / ou dessins, normes et règlements en vigueur, mentionnées dans le présent document

VWR International bestätigt, dass das Produkte allen zutreffenden Spezifikations- und/oder Zeichnungs- und/oder Standardanforderungen sowie den im Dokument aufgeführten gültigen Vorschriften entspricht.

Cat. No.	Description	Pk
732-2890	Deep 96-well plate, 1,2 ml	10
732-2891	Deep 96-well plate, 1,2 ml, sterile	10
732-2892	Deep 96-well plate, 2,2 ml	10
732-2893	Deep 96-well plate, 2,2 ml, sterile	10

Material: Polypropylene

Product Dimensions:

Length**: 127.76 mm**
Width**: 85.59 mm**
Height**: 43.94 mm**

For 732-2890 and 732-2891:

Well-depth: 24.63 mm
Diameter: 8.12 mm

For 732-2892 and 732-2893:

Well-depth: 39.62 mm
Diameter: 8.12 mm

MicroBiological testing:

Certified testing is performed lot by lot. These tests include Nuclease (Rnase, Dnase), Endotoxin (pyrogen) and ATP. Certificates are on file at the production facility.

Certificate of Conformity

Performance Testing:


Each manufacturing lot is sampled and tested in accordance with Standard Operating Procedures.

Sterility: Sterile plates are certified with a sterility assurance level of (SAL) 10^{-6}.


Other characteristics:

- These PP deep well plates feature double alignment corner notches.
- Labelled on three sides with an alphanumeric grid.
- Smooth inner well surfaces ensure that all of the sample can be pipetted.
- Designed for automation.
- Number on the alphanumeric grid alternate between standard face and bold face print to make the plate easier to work with.
- Can be sealed using film or mats.
- Freezable to –80 °C; autoclavable to 122 °C
- Certified free from endotoxin, RNase, DNase and ATP

732-2892



Certificate of Conformity



Disclaimer: *VWR wants to state that this declaration will not discharge the user from his obligation to ensure the product is suitable for the intended use. - If there is any doubt ask us.*

Date of issue: **February 2018**