Certificate of Conformity

VWR ZIP BAGS, SINGLE POCKET WITHOUT WRITING AREA

VWR International certifies that the specified products meet all VWR specifications and quality criteria for release.The products comply in every respect with the relevant particular specifications, drawings, relevant standards and regulations in force.

VWR International certifie que les produits mentionés respectent toutes les specifications et critères de qualité VWR. Les produits satisfont aux spécifications particulières, aux plans, aux normes et règlements en vigueur s'y rapportant.

VWR International bestätigt hiermit, daß die angegebenen Produkte alle VWR Spezifikationen und Qualitätskriterien zur Freigabe erfüllen. Die Produkte entsprechen in jeder Hinsicht den maßgeblichen Spezifikationen, den Zeichnungen sowie den relevanten Normen und gültigen Vorschriften.

VWR Catalog No: **129-0325**

Lot No.: **080310**

L x W : **200mm x 300mm**

Pack : **1000**

Material: **LDPE C150Y+LLDPE 218W (Virgin)**

QC Testing: **Representative production samples are collected and inspected in accordance with current applicable product specifications.**

Sterility : **Product labeled as sterile is gamma irradiated and dosimetrically released upon US Association for the Advancement of Medical Instrumentation (AAMI) recommended practices in effect at the time of validation. Products labeled sterile meets a minimum requirement of 10^{-6} SAL.**

This document has been produced electronically and is valid without a signature

Date of Issue: **September 2011**

VWR International bvba/sprl, Researchpark Haasrode 2020, Geldenaaksebaan 464, 3001 Leuven, Belgium
http://www.vwr.com