Certificate of Conformity

VWR VOLUMETRIC FLASKS, CLASS A, BLUE GRADUATION

VWR International certifies that the specified products meet all VWR specifications and quality criteria for release.The products comply in every respect with the relevant particular specifications, drawings, relevant standards and regulations in force.

VWR International certifie que les produits mentionés respectent toutes les specifications et critères de qualité VWR. Les produits satisfont aux spécifications particulières, aux plans, aux normes et règlements en vigueur s'y rapportant.

VWR International bestätigt hiermit, daß die angegebenen Produkte alle VWR Spezifikationen und Qualitätskriterien zur Freigabe erfüllen. Die Produkte entsprechen in jeder Hinsicht den maßgeblichen Spezifikationen, den Zeichnungen sowie den relevanten Normen und gültigen Vorschriften.

Cat. No.	Lot No.	Capacity ml	NS	Tolerance	Pack
612-3820	0412	2000	29/32	0.600 ± ml	1

Material : Borosilicate glass, with PE Stopper

QC Testing: Representative production samples are collected and inspected in accordance with current applicable product specifications.

This document has been produced electronically and is valid without a signature

Date of issue: **August 2012**