



Certificate of Conformity

VWR HDPE SQUARE BOTTLE 1L, FOR WATER SAMPLING, WITH 120mg/l $\text{Na}_2\text{S}_2\text{O}_3$

VWR International certifies that the specified products meet all VWR specifications and quality criteria for release. The products comply in every respect with the relevant particular specifications, drawings, relevant standards and regulations in force.

VWR International certifie que les produits mentionnés respectent toutes les spécifications et critères de qualité VWR. Les produits satisfont aux spécifications particulières, aux plans, aux normes et règlements en vigueur s'y rapportant.

VWR International bestätigt hiermit, daß die angegebenen Produkte alle VWR Spezifikationen und Qualitätskriterien zur Freigabe erfüllen. Die Produkte entsprechen in jeder Hinsicht den maßgeblichen Spezifikationen, den Zeichnungen sowie den relevanten Normen und gültigen Vorschriften.

VWR Catalog No: **331-0053**

Lot No: **473574**

Sterility Assurance Level: **Sterile**

Manufacturing Date (DD/MM/YYYY): **24/09/2012**

Expiry Date (MM/YY): **09/2014**

Dosage of Sodium Thiosulphate by gravimetry: **min. 120 mg/max. 140 mg**

This document has been produced electronically and is valid without a signature

Date of issue: **November 2012**



VWR International bvba/sprl, Researchpark Haasrode 2020, Geldenaaksebaan 464, 3001 Leuven, Belgium
<http://www.vwr.com>

Version: 2.0