Certificate of Conformity

VWR PP BOTTLES, TRANSLUCENT, WIDE MOUTH WITH SCREW CAP

VWR International certifies that the products specified conform to all specifications and quality criteria mentioned in this document.

VWR International certifie que les produits mentionnés respectent toutes les specifications et critères de qualité mentionnés dans ce document.

VWR International bestätigt hiermit, daß die angegebenen Produkte alle in diesem Dokument erwähnten Spezifikationen und Qualitätskriterien erfüllen.

Catalog No.	Capacity (ml)	Thread	Ø ext. (mm)	Height (mm)	Pk
215-1898	2000	100-415	119	240	1
215-1899	4000	100-415	163	298	1
215-3449	30	28-415	35	62.5	12
215-3449S	30	28-415	35	62.5	1.000
215-3450	60	28-415	38.5	85	12
215-3450S	60	28-415	38.5	85	1.000
215-3451	125	38-415	50.5	98	12
215-3451S	125	38-415	50.5	98	500
215-3452	250	43-415	61.5	131	12
215-3452S	250	43-415	61.5	131	250
215-3453	500	53-415	72.5	169	12
215-3453S	500	53-415	72.5	169	125
215-3454	1000	63-415	91.5	199	6
215-3454S	1000	63-415	91.5	199	50

Material: Polypropylene

Other Characteristics:

- Can be used anywhere in the laboratory, as well as for collecting samples
- Wide opening allows rapid filling and dispensing of powders, pastes and liquids
- Four litre bottles have integrated handle design
- Good chemical resistance
- PP heat resistant up to +135 °C, autoclavable
- Leakproof

Disclaimer: *VWR wants to state that this declaration will not discharge the user from his obligation to ensure the product is suitable for the intended use. - If there is any doubt ask us.*

Date of issue: **June 2015**