Certificate of Conformity

VWR ERLENMEYER FLASKS WITH BEADED RIM

VWR International certifies that the products specified conform to all specifications and quality criteria mentioned in this document.

VWR International certifie que les produits mentionnés respectent toutes les specifications et critères de qualité mentionnés dans ce document.

VWR International bestätigt hiermit, daß die angegebenen Produkte alle in diesem Dokument erwähnten Spezifikationen und Qualitätskriterien erfüllen.

Erlenmeyer, narrow neck: **Erlenmeyer, wide neck:**

VWR Cat. N°	Capacity (ml)	VWR Cat. N°	Capacity (ml)
214-0157	200	214-0158	200
214-0248	25	214-1170	50
214-1130	50	214-1171	100
214-1131	100	214-1172	250
214-1132	250	214-1173	300
214-1133	500	214-1174	500
214-1134	1.000	214-1175	1.000
214-1135	2.000	214-1176	2.000
214-1136	3.000		
214-1137	5.000		

Intended uses: Laboratory applications

Material: Borosilicate 3.3 glass

Shape: Narrow or wide neck, with rim and graduation

Compliant to: ISO 3585 Borosilicate glass 3.3 - properties
USP (United States Pharmacopoeia) 23, section 661: Type 1
ASTM E 428-29 (1996): Type 1, class A
European Pharmacopoeia Paragraph 3.2.1

The compliance of demands is proved by regular tests and followed noticed tests from an independent laboratory:

- Test acc. to USP 23 (661) Containers; Section Physical Tests of Containers and EP 3.2.1
- Powered glass test and water attack at 121°C
- Test for arsenic (211): Photometrical detection
- Test acc. to ASTM E 438-92 (1996); Glasses

Certificate of Conformity

Chemical Properties: Tested acc. to USP 23 (661) Containers; Section Physical Tests of Containers and EP 3.2.1
Hydrolytic resistance at 98°C ISO 719-HGB 1
Hydrolytic resistance at 121°C ISO720-HGA 1
Acid resistance acc. to ISO 1776
Resistance to attack by a boiling aqueous solution of mixed alkali acc. to ISO 695-A2

Disclaimer: *VWR wants to state that this declaration will not discharge the user from his obligation to ensure the product is suitable for the intended use. - If there is any doubt ask us.*

Date of issue: **September 2015**